



## CERTIFICATE OF ANALYSIS 181672

### Client Details

<b>Client</b>	Walker Quarries Pty Ltd
<b>Attention</b>	Ray sharwood
<b>Address</b>	Lot 6, Great Western Hwy, Wallerawang, NSW

### Sample Details

<b>Your Reference</b>	<u>Walker Quarries</u>
<b>Number of Samples</b>	4 Dust Guage
<b>Date samples received</b>	08/12/2017
<b>Date completed instructions received</b>	08/12/2017

### Analysis Details

Please refer to the following pages for results, methodology summary and quality control data.  
Samples were analysed as received from the client. Results relate specifically to the samples as received.  
Results are reported on a dry weight basis for solids and on an as received basis for other matrices.

### Report Details

<b>Date results requested by</b>	15/12/2017
<b>Date of Issue</b>	15/12/2017
NATA Accreditation Number 2901. This document shall not be reproduced except in full.	
Accredited for compliance with ISO/IEC 17025 - Testing. <b>Tests not covered by NATA are denoted with *</b>	

#### Results Approved By

Nick Sarlamis, Inorganics Supervisor

#### Authorised By

David Springer, General Manager

Client Reference: Walker Quarries

Dust Deposition AS 3580.10.1					
Our Reference		181672-1	181672-2	181672-3	181672-4
Your Reference	UNITS	D1	D2	D3	D4
Type of sample		Dust Guage	Dust Guage	Dust Guage	Dust Guage
Dust Gauge Start Date	--	06/11/2017	06/11/2017	06/11/2017	06/11/2017
Dust Gauge End Date	--	04/12/2017	04/12/2017	04/12/2017	04/12/2017
Dust - No. of Days Collected	--	29	29	29	29
Insoluble Solids	g/m <sup>2</sup> /month	0.6	1	0.5	0.3
Combustible Matter	g/m <sup>2</sup> /month	0.3805	0.7785	0.1405	0.1756
Ash	g/m <sup>2</sup> /month	0.1756	0.2107	0.3746	0.1580

**Client Reference: Walker Quarries**

Method ID	Methodology Summary
<b>AS 3580.10.1</b>	Analysis of dust gauges by AS 3580.10.1 (analysis only).

## Result Definitions

<b>NT</b>	Not tested
<b>NA</b>	Test not required
<b>INS</b>	Insufficient sample for this test
<b>PQL</b>	Practical Quantitation Limit
<b>&lt;</b>	Less than
<b>&gt;</b>	Greater than
<b>RPD</b>	Relative Percent Difference
<b>LCS</b>	Laboratory Control Sample
<b>NS</b>	Not specified
<b>NEPM</b>	National Environmental Protection Measure
<b>NR</b>	Not Reported

## Quality Control Definitions

<b>Blank</b>	This is the component of the analytical signal which is not derived from the sample but from reagents, glassware etc, can be determined by processing solvents and reagents in exactly the same manner as for samples.
<b>Duplicate</b>	This is the complete duplicate analysis of a sample from the process batch. If possible, the sample selected should be one where the analyte concentration is easily measurable.
<b>Matrix Spike</b>	A portion of the sample is spiked with a known concentration of target analyte. The purpose of the matrix spike is to monitor the performance of the analytical method used and to determine whether matrix interferences exist.
<b>LCS (Laboratory Control Sample)</b>	This comprises either a standard reference material or a control matrix (such as a blank sand or water) fortified with analytes representative of the analyte class. It is simply a check sample.
<b>Surrogate Spike</b>	Surrogates are known additions to each sample, blank, matrix spike and LCS in a batch, of compounds which are similar to the analyte of interest, however are not expected to be found in real samples.

Australian Drinking Water Guidelines recommend that Thermotolerant Coliform, Faecal Enterococci, & E.Coli levels are less than 1cfu/100mL. The recommended maximums are taken from "Australian Drinking Water Guidelines", published by NHMRC & ARMC 2011.

## Laboratory Acceptance Criteria

Duplicate sample and matrix spike recoveries may not be reported on smaller jobs, however, were analysed at a frequency to meet or exceed NEPM requirements. All samples are tested in batches of 20. The duplicate sample RPD and matrix spike recoveries for the batch were within the laboratory acceptance criteria.

Filters, swabs, wipes, tubes and badges will not have duplicate data as the whole sample is generally extracted during sample extraction.

Spikes for Physical and Aggregate Tests are not applicable.

For VOCs in water samples, three vials are required for duplicate or spike analysis.

Duplicates: <5xPQL - any RPD is acceptable; >5xPQL - 0-50% RPD is acceptable.

Matrix Spikes, LCS and Surrogate recoveries: Generally 70-130% for inorganics/metals; 60-140% for organics (+/-50% surrogates) and 10-140% for labile SVOCs (including labile surrogates), ultra trace organics and speciated phenols is acceptable.

In circumstances where no duplicate and/or sample spike has been reported at 1 in 10 and/or 1 in 20 samples respectively, the sample volume submitted was insufficient in order to satisfy laboratory QA/QC protocols.

When samples are received where certain analytes are outside of recommended technical holding times (THTs), the analysis has proceeded. Where analytes are on the verge of breaching THTs, every effort will be made to analyse within the THT or as soon as practicable.

Where sampling dates are not provided, Envirolab are not in a position to comment on the validity of the analysis where recommended technical holding times may have been breached.

Measurement Uncertainty estimates are available for most tests upon request.