



Envirolab Services Pty Ltd

ABN 37 112 535 645

12 Ashley St Chatswood NSW 2067

ph 02 9910 6200 fax 02 9910 6201

customerservice@envirolab.com.au

www.envirolab.com.au

CERTIFICATE OF ANALYSIS 183263

Client Details

Client	Walker Quarries
Attention	Ray Sharwood
Address	Lot 6 Great Western Highway, Wallerawang, NSW, 2845

Sample Details

Your Reference	<u>Walker Quarries</u>
Number of Samples	
Date samples received	15/01/2018
Date completed instructions received	15/01/2018

Analysis Details

Please refer to the following pages for results, methodology summary and quality control data.

Samples were analysed as received from the client. Results relate specifically to the samples as received.

Results are reported on a dry weight basis for solids and on an as received basis for other matrices.

Report Details

Date results requested by 22/01/2018

Date of Issue 22/01/2018

NATA Accreditation Number 2901. This document shall not be reproduced except in full.

Accredited for compliance with ISO/IEC 17025 - Testing. **Tests not covered by NATA are denoted with ***

Results Approved By

Priya Samarawickrama, Senior Chemist

Authorised By

David Springer, General Manager

Client Reference: Walker Quarries

Dust Deposition AS 3580.10.1					
Our Reference		183263-1	183263-2	183263-3	183263-4
Your Reference	UNITS	DGS-1	DGS-2	DGS-3	DGS-4
Date Sampled		1/12/17-3/1/2018	1/12/17-3/1/2018	1/12/17-3/1/2018	1/12/17-3/1/2018
Dust Gauge Start Date	--	01/12/2017	01/12/2017	01/12/2017	01/12/2017
Dust Gauge End Date	--	03/01/2018	03/01/2018	03/01/2018	03/01/2018
Dust - No. of Days Collected	--	34	34	34	34
Date prepared	-	16/01/2018	16/01/2018	16/01/2018	16/01/2018
Date analysed	-	16/01/2018	16/01/2018	16/01/2018	16/01/2018
Insoluble Solids	g/m ² /month	0.7	9.8	<0.1	0.9
Combustible Matter	g/m ² /month	0.3944	7.7583	0.0349	0.4943
Ash	g/m ² /month	0.3345	2.0269	0.0200	0.3944
Soluble Matter	g/m ² /month	<0.1	2.4	<0.1	0.1
Total Solids	g/m ² /month	0.7	12	0.1	0.99

Client Reference: Walker Quarries

Method ID	Methodology Summary
AS 3580.10.1	Analysis of dust gauges by AS 3580.10.1 (analysis only).

Result Definitions

NT	Not tested
NA	Test not required
INS	Insufficient sample for this test
PQL	Practical Quantitation Limit
<	Less than
>	Greater than
RPD	Relative Percent Difference
LCS	Laboratory Control Sample
NS	Not specified
NEPM	National Environmental Protection Measure
NR	Not Reported

Quality Control Definitions

Blank	This is the component of the analytical signal which is not derived from the sample but from reagents, glassware etc, can be determined by processing solvents and reagents in exactly the same manner as for samples.
Duplicate	This is the complete duplicate analysis of a sample from the process batch. If possible, the sample selected should be one where the analyte concentration is easily measurable.
Matrix Spike	A portion of the sample is spiked with a known concentration of target analyte. The purpose of the matrix spike is to monitor the performance of the analytical method used and to determine whether matrix interferences exist.
LCS (Laboratory Control Sample)	This comprises either a standard reference material or a control matrix (such as a blank sand or water) fortified with analytes representative of the analyte class. It is simply a check sample.
Surrogate Spike	Surrogates are known additions to each sample, blank, matrix spike and LCS in a batch, of compounds which are similar to the analyte of interest, however are not expected to be found in real samples.

Australian Drinking Water Guidelines recommend that Thermotolerant Coliform, Faecal Enterococci, & E.Coli levels are less than 1cfu/100mL. The recommended maximums are taken from "Australian Drinking Water Guidelines", published by NHMRC & ARMC 2011.

Laboratory Acceptance Criteria

Duplicate sample and matrix spike recoveries may not be reported on smaller jobs, however, were analysed at a frequency to meet or exceed NEPM requirements. All samples are tested in batches of 20. The duplicate sample RPD and matrix spike recoveries for the batch were within the laboratory acceptance criteria.

Filters, swabs, wipes, tubes and badges will not have duplicate data as the whole sample is generally extracted during sample extraction.

Spikes for Physical and Aggregate Tests are not applicable.

For VOCs in water samples, three vials are required for duplicate or spike analysis.

Duplicates: <5xPQL - any RPD is acceptable; >5xPQL - 0-50% RPD is acceptable.

Matrix Spikes, LCS and Surrogate recoveries: Generally 70-130% for inorganics/metals; 60-140% for organics (+/-50% surrogates) and 10-140% for labile SVOCs (including labile surrogates), ultra trace organics and speciated phenols is acceptable.

In circumstances where no duplicate and/or sample spike has been reported at 1 in 10 and/or 1 in 20 samples respectively, the sample volume submitted was insufficient in order to satisfy laboratory QA/QC protocols.

When samples are received where certain analytes are outside of recommended technical holding times (THTs), the analysis has proceeded. Where analytes are on the verge of breaching THTs, every effort will be made to analyse within the THT or as soon as practicable.

Where sampling dates are not provided, Envirolab are not in a position to comment on the validity of the analysis where recommended technical holding times may have been breached.

Measurement Uncertainty estimates are available for most tests upon request.